

AUG 15 2000

**510(k) Summary**

**PRYM MEDICAL Ltd.**

**Sansibar™ CPAP System**

**510(k) Number K 000888**

**Submitter's Name:**

PRYM MEDICAL Ltd.  
12 Hamada Street  
Beit Hatamar  
Rehovot 76703, Israel  
Tel: 972-8-931-6450  
Fax: 972-8-931-6470

**Contact Person:**

Shoshana Friedman, RAC  
Push-Med Ltd.  
117, Ahuza St., Ra'ananna 43373, Israel  
Tel: 972-9-7718130  
Fax: 972-9-7718131

**Trade Name:**

Sansibar™ CPAP System

**Classification Name:**

Ventilator, Non-Continuous

**Classification:**

Non-continuous ventilators are II (Product Code BZD).

**Predicate Device:**

The Sansibar™ CPAP System is substantially equivalent to the Aria CPAP System (Respironics, Inc.) Cleared under K953341.

**Performance Standards:**

No performance standards have been established for such devices under sections 514 of the Federal Food, Drug, and Cosmetics Act. The Sansibar™ CPAP System has been tested for compliance with standards IEC 60601-1 (and amendments), IEC 60601-1-1 (and amendments) and IEC 60601-1-2.

**Indication for use:**

The Sansibar™ CPAP System is indicated for the treatment of adult Obstructive Sleep Apnea (OSA).

**Device Description:**

- The Sansibar™ CPAP System is used to assist with patient breathing while sleeping, for the purpose of treating Obstructive Sleep Apnea (OSA). This is by the delivery of Continuous Positive Airway Pressure between 4 and 20 cmH<sub>2</sub>O to prevent airway obstruction.

The Sansibar™ CPAP System is powered by input voltage 85 to 265 VAC, 50/60Hz. The blower motor provides positive air pressure, which can be adjusted as prescribed by a physician. The blower speed is directly related to air pressure, and is controlled by software. An air filter is located at the back of the device to keep the air clean of dust and other small particles.

The Sansibar™ CPAP System has two main user accessible controls, on/off and ramp time, and three multifunctional control keys for menu navigation and user settings. The pressure setting is set by prescription and is locked-out from the end user's reach. The device is designed for use with a remote control and a remote signal interface. It uses a microcontroller to set the various control parameters, turn features on and off, and view the set parameters.

**Substantial Equivalence:**

Based on a series of safety and performance testing, Prym Medical Ltd. believes that the Sansibar™ CPAP System is substantially equivalent to the predicate device cited above without raising new safety and/or effectiveness issues.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

AUG 15 2000

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Prym Medical, Ltd.  
c/o Ms. Shoshana Friedman  
Push-Med Ltd.  
117, Ahuza Street, Ra'ananna 43373  
ISRAEL

Re: K000888  
Sansibar™ CPAP System  
Product Code: 73 BZD  
Dated: June 4, 2000  
Received: July 6, 2000

Dear Ms. Friedman:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 - Ms. Shoshana Friedman

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4648. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597, or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,

*for Mark N. Melhus*

James E. Dillard III  
Director  
Division of Cardiovascular and  
Respiratory Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## INDICATIONS FOR USE STATEMENT

510(k) Number (if known): K000888

Device Name: Sansibar™ CPAP System

### Indications for Use:

The Sansibar™ CPAP System is indicated for the treatment of adult Obstructive Sleep Apnea (OSA).

for Mark N. Melker  
(Division Sign-Off)  
Division of Cardiovascular, Respiratory,  
and Neurological Devices  
510(k) Number K000888

(PLEASE DO NOT WRITE BELOW THIS LINE -CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use ✓  
(Per 21 CFR 801.109)

OR

Over the Counter Use \_\_\_\_\_